



BSL-2 laboratory verification routine

Laboratorio de Genómica Viral y Humana BSL-3, Facultad de Medicina UASLP

Last updated May 02, 2024 v4



Use ✓ if ok, or ■ if not.

Year: Month:

Day:

S M T W T F S

Warehouse Refrigerator between -4 and +4°C.

Aisle Primedic Defi-B defibrillator operative.

RT-PCR Refrigerator between -4 and +4°C.
Instruments and air conditioning unit are off.


Cell Freezer between -25 and -40°C.
Biology Refrigerator between -4 and +4°C.
Instruments and air conditioning unit are off.

Control room Air handling unit A or B is on.
Air handling unit is < 60 Hz and below 11 amps.
Four extractors are on.
CCTV screen is on and 14 cameras visible.
Server rack is on (flashing red/green LEDs).
Electrical room UPS is on (green LED).
Electrical room variator is < 60 Hz.

Molecular Biology Air conditioning unit ON and set to 25 °C.
Refrigerator between -4 and +4°C.
Freezer between -17 and -30°C.
Freezer alarm on and ok (green LED).
Liquid nitrogen level ok (green) in Tank 1.
Liquid nitrogen level ok (green) in Tank 2.
Main ultra-low freezer is ON and between -60 and -80°C.
Backup ultra-low freezer is ON and between -60 and -80°C.
Main or Backup ultra-low freezers are OFF and free of mold.
Emergency powerplant on terrace and covered with plastic.
Two CO₂ cylinders are present on the terrace.
Roof-top access staircase padlocked.
Main sink faucet is left open for 30 seconds.
Main sink eyewash station is left open for 30 seconds.
Applied 7500 cycler UPS NOT alarmed NOR reset.

Wash-up room Semaphore in green status.
Digital manometer is in green status.
Pass-through indicator is in green status.
Access door to wash-up room is fully closed and locked.

Comments:



Checked by